

PRODUCT BULLETIN

Product(s):	Fabius GS
Bulletin Number:	160
Release Date:	November 25, 2003
Pages:	2

Subject: Fabius GS COSY breathing system change


Effective immediately, all new production Fabius GS systems are being shipped with a COSY that is compatible with the CLIC absorber system.

General Overview


The COSY change is a resolution of the compatibility issue with the CLIC absorber system. As a result of this change, the part number for the COSY has changed.

	<u>US</u>	<u>Canada</u>
New COSY Part #	4118378	4118379

With the new COSY, the bag location has moved slightly.



Old COSY



New COSY

As a result of the bag location change, the new COSY will require a different bag arm mount.

Swivel Bag Arm Part # (new COSY only)	4118380	\$ 450
--	----------------	---------------

Please note that the old COSY is only compatible with the old bag arm mount, and the new COSY is only compatible with the new bag arm mount.

Rob Clark has updated the quote templates on the Anesthesia Marketing Library to reflect this change.

Field Install Plan

For existing Fabius GS customers who are currently using the CLIC, or who originally ordered the CLIC, an upgrade program to the new COSY will be established. The details of this program are not yet finalized, and due to inventory restrictions will not be implemented until the beginning of 2004.

Feel free to contact Mary Kyryk at ext. 2619 or kyrykm@draegermed.com if you have additional questions.